



測定値 (mg/L)	1回目	10.3	11.5	13.3	12.4	11.8	12.9	15.6	12.9	12.9	13.5	8.2	11.8	11.0	9.0	9.9	11.5	11.0	12.0	11.6	9.0	8.8	9.6	10.5	10.3	10.5	11.0	9.9	12.2	13.3	11.9	11.5	10.3	11.9	
	2回目	10.4	10.7	13.4	12.3	12.0	13.1	15.5	12.8	12.8	13.4	7.0	12.3	11.0	9.7	10.3	11.4	11.1	12.2	11.8	9.5	9.7	10.7	11.2	10.8	9.2	11.2	10.2	12.8	13.2	11.6	11.5	10.6	11.9	
	3回目	10.3	10.5	13.5	12.1	11.9	12.9	15.0	13.2	12.7	13.6	7.2	11.2	10.4	11.3	10.5	11.6	11.1	12.0	11.7	9.8	9.5	11.3	11.1	9.7	10.7	11.2	9.7	12.7	13.5	12.0	11.2	9.9	12.0	
平均値(mg/L)①	10.3	10.9	13.4	12.3	11.9	13.0	15.4	13.0	12.8	13.5	7.5	11.8	10.8	10.0	10.2	11.5	11.1	12.1	11.7	9.4	9.3	10.5	10.9	10.3	10.1	11.1	9.9	12.6	13.3	11.8	11.4	10.3	11.9		
標準偏差②	0.06	0.53	0.10	0.15	0.10	0.12	0.32	0.21	0.10	0.10	0.64	0.55	0.35	1.18	0.31	0.10	0.06	0.12	0.10	0.40	0.47	0.86	0.38	0.55	0.81	0.12	0.25	0.32	0.15	0.21	0.17	0.35	0.06		
CV(室内)(%) ((2)÷(1))	0.56	4.86	0.75	1.25	0.84	0.89	2.09	1.61	0.78	0.74	8.61	4.68	3.21	11.79	2.99	0.87	0.52	0.96	0.86	4.29	5.07	8.19	3.46	5.36	8.04	1.04	2.54	2.56	1.15	1.76	1.52	3.42	0.48		
参加者平均との差	-1.06	-0.49	2.01	0.88	0.51	1.58	3.98	1.58	1.41	2.11	-3.92	0.38	-0.59	-1.39	-1.16	0.11	-0.32	0.68	0.31	-1.96	-2.06	-0.86	-0.46	-1.12	-1.26	-0.26	-1.46	1.18	1.94	0.44	0.01	-1.12	0.54		
Zスコア	-0.69	-0.32	1.32	0.58	0.33	1.04	2.61	1.04	0.92	1.38	-2.57	0.25	-0.39	-0.91	-0.76	0.07	-0.21	0.45	0.20	-1.28	-1.35	-0.56	-0.30	-0.73	-0.83	-0.17	-0.96	0.77	1.27	0.29	0.01	-0.73	0.35		
Grubbs検定による棄却																																			

平均値	11.4
標準偏差	1.53
CV(室内)%	2.96

Zスコア	事業所数	33
$ Z \leq 2$		31
$2 < Z < 3$		2
$3 \leq Z $		0