



測定値 (mg/L)	1回目	23.6	20.0	20.4	26.5	21.8	20.6	22.2	24.5	22.2	20.4	20.9	18.0	20.8	20.8	20.7	20.2	19.8	15.0	17.9	21.4	18.8	23.2	20.5	22.3	22.1	22.6	21.4	22.4	22.6	23.0
	2回目	23.6	20.2	20.5	26.8	21.7	20.6	21.9	24.7	20.2	20.1	21.4	17.0	19.2	20.2	20.5	19.2	20.9	15.0	17.8	21.5	19.5	23.3	21.4	22.0	21.3	22.4	21.0	23.8	22.5	23.8
	3回目	23.5	20.0	20.6	26.1	21.7	20.7	21.4	25.2	19.0	19.8	20.5	17.8	21.3	20.3	20.4	19.4	21.3	14.9	17.4	21.5	20.9	22.9	21.8	22.3	21.8	22.7	21.0	20.7	22.7	22.8
平均値(mg/L)①		23.6	20.1	20.5	26.5	21.7	20.6	21.8	24.8	20.5	20.1	20.9	17.6	20.4	20.4	20.5	19.6	20.7	15.0	17.7	21.5	19.7	23.1	21.2	22.2	21.7	22.6	21.1	22.3	22.6	23.2
標準偏差②		0.06	0.12	0.10	0.35	0.06	0.06	0.40	0.36	1.62	0.30	0.45	0.53	1.10	0.32	0.15	0.53	0.78	0.06	0.26	0.06	1.07	0.21	0.67	0.17	0.40	0.15	0.23	1.55	0.10	0.53
CV(室内)(%) (②/①)		0.25	0.58	0.49	1.33	0.27	0.28	1.85	1.45	7.90	1.49	2.15	3.01	5.37	1.57	0.74	2.70	3.76	0.39	1.50	0.27	5.42	0.90	3.14	0.78	1.86	0.68	1.09	6.96	0.44	2.28
参加者平均との差		2.43	-1.07	-0.64	5.33	0.59	-0.51	0.69	3.66	-0.67	-1.04	-0.21	-3.54	-0.71	-0.71	-0.61	-1.54	-0.47	-6.17	-3.44	0.33	-1.41	1.99	0.09	1.06	0.59	1.43	-0.01	1.16	1.46	2.06
Zスコア		1.12	-0.49	-0.30	2.46	0.27	-0.24	0.32	1.69	-0.31	-0.48	-0.10	-1.64	-0.33	-0.33	-0.28	-0.71	-0.22	-2.85	-1.59	0.15	-0.65	0.92	0.04	0.49	0.27	0.66	0.00	0.54	0.67	0.95
Grubss検定による棄却																															

平均値	21.1
標準偏差	2.16
CV(室内)%	2.03

Zスコア	事業所数	30
$ Z \leq 2$	28	
$2 < Z < 3$	2	
$3 \leq Z $	0	